


1 June 2023

Addendum 4 for RFP 23-34 – Powder X-ray Diffractometer System

This addendum updates an incorrect bullet point in the provided bid specifications.



Amber Floyd

Buyer, Procurement, and Contract Services

Update regarding section 2.2.2 “General Equipment Description” for BID No. 23-34. Bullet point #16, “Must include a diffracted beam graphite monochromator for maximum intensity” This should state, “The monochromator must be graphite or multilayer, no nickel foil will be accepted.”